

Instruction Manual

HI 3882 pH 3.0-5.0 Test Kit

HANNA
instruments
www.hannainst.com

Dear Customer,
Thank you for choosing a Hanna Product.

Please read the instructions carefully before using the chemical test kit. It will provide you with the necessary information for correct use of the kit.

Remove the chemical test kit from the packing material and examine it carefully to make sure that no damage has occurred during shipping. If there is any noticeable damage, notify your Dealer or the nearest Hanna office immediately.

Each kit is supplied with:

- HI 3882-0 pH 3.0-5.0 Reagent, 1 bottle with dropper (25 mL);
- 1 checker disc;
- 2 glass vials with caps;
- 1 plastic pipette (3 mL).

Note: Any damaged or defective item must be returned in its original packing materials.

SPECIFICATIONS

Range	3.0 to 5.0 as pH unit
Smallest Increment	0.1 as pH unit
Analysis Method	Colorimetric
Sample Size	5 mL
Number of Tests	200
Case Dimensions	165x150x38 mm (6.5x5.9x1.5")
Shipping Weight	215 g (7.6 oz.)

SIGNIFICANCE AND USE

pH represents acidity or alkalinity of an aqueous solution and is proportional to the hydrogen-ion concentration of the solution. Under neutral conditions water is dissociated into the OH^- and H^+ ions in equal ratio and hence it has a pH of 7. When bases or acids are added to a water solution they ionize, increasing the concentration of OH^- or H^+ , respectively. Thus solutions with a pH of 1-3 contain strong acids, whereas those with a pH of 4-6 contain weak acids. Weak bases result in solutions of pH 8-10 and strong bases in pH of 11-13.

Examples of pH value for some liquids:

Liquid	pH Value
Sea water	7.8-8.2
Gastric juices	1.7
Milk	6.5-7
Soil	6-7 (optimum for crops)

CHEMICAL REACTION


HI 3882-0 reagent reacts in contact with the aqueous solution changing its color according to the hydrogen-ion concentration (pH) in the given range.

INSTRUCTIONS


READ THE ENTIRE INSTRUCTIONS BEFORE USING THE KIT

Note: Always shake the reagent bottle before use.

- Using the plastic pipette, fill each glass vial with 5 mL of sample, up to the mark.
- Insert one of them into the left hand opening of the checker disc. This is the




- Add to the other glass vial 2 drops of HI 3882-0 reagent. Replace the cap and mix the solution by gently swirling. This is the reacted sample.




- Remove the cap and insert the reacted sample into the right hand opening of the checker disc.

- Hold the checker disc so that a light source illuminates the samples from the back windows.



- Rotate the disc while looking at the color test windows and stop when you find the color match. Read the value in the result window and record it in pH units.



Note: A concentration of Chlorine above 50 ppm causes interference inhibiting the development of the color. In this case require the HI 3882/0 Test Kit.

For best results: Intensely colored samples will make the color matching determination difficult and they should be adequately treated before performing the test. Suspended matter in large amounts should be removed by prior filtration.

Caution: Ultraviolet radiation may cause fading of colors. When not in use, keep the disc protected from light, in a cool and dry place.

REFERENCES

Vogel's, Quantitative Chemical Analysis, 5th Ed., Longman Scientific & Technical.

HEALTH AND SAFETY

The chemicals contained in this kit may be hazardous if improperly handled. Read Health and Safety Data Sheet before performing this test.