

Reagents for Photometers

Code	Description	Code	Description
Alkalinity, colorimetric method			
Range:	0 to 500 mg/L	Chlorine, free and total HR, DPD method	
Reagents		Range:	0.00 to 10.00 mg/L
HI 93755-01	Reagent kit for 100 tests (CaCO ₃)	Reagents	
HI 93755-03	Reagent kit for 300 tests (CaCO ₃)	HI 93734-01	Powder reagent kit for 100 tests (Cl ₂ free or total)
Aluminum, aluminon method			
Range:	0.00 to 1.00 mg/L	HI 93734-03	Powder reagent kit for 300 tests (Cl ₂ free or total)
Reagents		Chlorine, free, ULR, DPD method	
HI 93712-01	Reagent kit for 100 tests (Al)	Range:	0.000 to 0.500 mg/L
HI 93712-03	Reagent kit for 300 tests (Al)	Reagents	
Ammonia HR, Nessler method			
Range:	0 to 100 mg/L	HI 95762-01	Reagent kit for 100 tests (free Cl ₂)
Reagents		HI 95762-03	Reagent kit for 300 tests (free Cl ₂)
HI 93764B-25	Reagent kit for 25 test vials (N-NH ₃ HR)	Chlorine, free, DPD method	
Range:	0.0 to 50.0 mg/L	Range:	0.00 to 2.50 mg/L (5.00 mg/L)****
Reagents		Reagents	
HI 93733-01	Reagent kit for 100 tests (NH ₄ ⁺ HR)	HI 93701-01	Powder reagent kit for 100 tests (free Cl ₂)
HI 93733-03	Reagent kit for 300 tests (NH ₄ ⁺ HR)	HI 93701-03	Powder reagent kit for 300 tests (free Cl ₂)
Ammonia LR, Nessler method			
Range:	0.00 to 3.00 mg/L	HI 93701-F	Liquid reagent Kit for 300 tests (free Cl ₂)
Reagents		Chlorine, total, ULR, DPD method	
HI 93700-01	Reagent kit for 100 tests (NH ₃ -N LR)	Range:	0.000 to 0.500 mg/L
HI 93700-03	Reagent kit for 300 tests (NH ₃ -N LR)	Reagents	
HI 93764A-25	Reagent kit for 25 test vials (NH ₃ -N LR)	HI 95761-01	Reagent kit for 100 tests (total Cl ₂)
Ammonia MR, Nessler method			
Range:	0.00 to 9.99 mg/L (10.00 mg/L)****	HI 95761-03	Reagent kit for 300 tests (total Cl ₂)
Reagents		Chlorine, total, DPD method	
HI 93715-01	Reagent kit for 100 tests (NH ₃ -N MR)	Range:	0.00 to 3.50 mg/L (5.00 mg/L)****
HI 93715-03	Reagent kit for 300 tests (NH ₃ -N MR)	Reagents	
Bromine, DPD method			
Range:	0.00 to 8.00 mg/L (10.00 mg/L)****	HI 93711-01	Powder reagent kit for 100 tests (total Cl ₂)
Reagents		HI 93711-03	Powder reagent kit for 300 tests (total Cl ₂)
HI 93716-01	Reagent kit for 100 tests (Br ₂)	HI 93701-T	Liquid reagent Kit for 300 tests (total Cl ₂)
HI 93716-03	Reagent kit for 300 tests (Br ₂)	HI 93711-D3	DPD3 reagent for 600 tests (total Cl ₂)
Calcium and magnesium HR, oxalate (Ca) and calmagite (Mg) methods			
Range:	0 to 400 mg/L (Ca) 0 to 150 mg/L (Mg)	Chromium VI HR, diphenylcarbohydrazide method	
Reagents		Range:	0 to 1000 µg/L
HI 93752-01	Reagent kit for 100 tests (50 Ca and 50 Mg)	Reagents	
HI 93752-03	Reagent kit for 300 tests (150 Ca and 150 Mg)	HI 93723-01	Reagent kit for 100 tests (Cr VI HR)
Chloride, mercury (II) thiocyanate method			
Range:	0.0 to 20.0 mg/L	HI 93723-03	Reagent kit for 300 tests (Cr VI HR)
Reagents		Chromium VI LR, diphenylcarbohydrazide method	
HI 93753-01	Reagent kit for 100 tests (Cl ⁻)	Range:	0 to 300 µg/L
HI 93753-03	Reagent kit for 300 tests (Cl ⁻)	Reagents	
Chlorine Dioxide, chlorophenol red method			
Range:	0.00 to 2.00 mg/L	HI 93749-01	Reagent kit for 100 tests (Cr VI LR)
Reagents		HI 93749-03	Reagent kit for 300 tests (Cr VI LR)
HI 93738-01	Reagent kit for 100 tests (ClO ₂)	COD, HR, dichromate method	
HI 93738-03	Reagent kit for 300 tests (ClO ₂)	Range:	0 to 150 mg/L
Chlorine, ULR, iodometric method			
Range:	0 to 500 mg/L	Reagents	
Reagents		HI 93754C-25	Reagent kit for 25 test vials (O ₂)
HI 95711-01	Reagent kit for 100 tests (Cl ₂)	COD, LR EPA*, dichromate method	
HI 95711-03	Reagent kit for 300 tests (Cl ₂)	Range:	0 to 150 mg/L
Chlorine, free, ULR, DPD method			
Range:	0.000 to 0.500 mg/L	Reagents	
Reagents		HI 93754A-25	Reagent kit for 25 test vials (O ₂)
HI 95762-01	Reagent kit for 100 tests (free Cl ₂)	COD, LR ISO**, dichromate method	
HI 95762-03	Reagent kit for 300 tests (free Cl ₂)	Range:	0 to 150 mg/L
Chlorine, free, DPD method			
Range:	0.00 to 2.50 mg/L (5.00 mg/L)****	Reagents	
Reagents		HI 93754F-25	Reagent kit for 25 test vials (O ₂)
HI 93701-01	Powder reagent kit for 100 tests (free Cl ₂)	COD, LR, dichromate method mercury free***	
HI 93701-03	Powder reagent kit for 300 tests (free Cl ₂)	Range:	0 to 15000 mg/L
HI 93701-F	Liquid reagent Kit for 300 tests (free Cl ₂)	Reagents	
Chlorine, total, ULR, DPD method			
Range:	0.000 to 0.500 mg/L	HI 93754D-25	Reagent kit for 25 test vials (O ₂)
Reagents			
HI 95761-01	Reagent kit for 100 tests (total Cl ₂)		
HI 95761-03	Reagent kit for 300 tests (total Cl ₂)		

* The sulphuric acid method is officially recognized by EPA for wastewater.
 ** HI 93754F-25 and HI 93754G-25 are compliant to ISO 15705 method.

*** It is a recommended method for generic analysis and if chlorine is absent.
 **** Some meters allow the user to extend the range to the value between brackets.

Reagents for Photometers

Code	Description
COD, MR, dichromate method mercury free***	
Range:	0 to 1500 mg/L
Reagents	
HI 93754E-25	Reagent kit for 25 test vials (O ₂)
COD, MR EPA*, dichromate method	
Range:	0 to 1500 mg/L
Reagents	
HI 93754B-25	Reagent kit for 25 test vials (O ₂)
COD, MR ISO**, dichromate method	
Range:	0 to 1000 mg/L
Reagents	
HI 93754G-25	Reagent kit for 25 test vials (O ₂)
Copper HR, bichinonate acid method	
Range:	0.00 to 5.00 mg/L
Reagents	
HI 93702-01	Reagent kit for 100 tests (Cu HR)
HI 93702-03	Reagent kit for 300 tests (Cu HR)
HI 93702T-01	Reagent kit for total copper 100 tests (Cu HR)
HI 93702T-03	Reagent kit for total copper 300 tests (Cu HR)
Copper LR, bichinonate acid method	
Range:	0.000 to 1.500 mg/L
Reagents	
HI 95747-01	Reagent kit for 100 tests (Cu LR)
HI 95747-03	Reagent kit for 300 tests (Cu LR)
Cyanide, pyridine-pyrazolone method	
Range:	0.000 to 0.200 mg/L
Reagents	
HI 93714-01	Reagent kit for 100 tests (CN ⁻)
HI 93714-03	Reagent kit for 300 tests (CN ⁻)
Cyanuric Acid, turbidimetric method	
Range:	0 to 80 mg/L (200 mg/L)****
Reagents	
HI 93722-01	Reagent kit for 100 tests (cyanuric acid)
HI 93722-03	Reagent kit for 300 tests (cyanuric acid)
Detergents, Anionic, MBAS method	
Range:	0.00 to 3.50 mg/L
Reagents	
HI 95769-01	Reagent kit for 40 tests (SDBS)
Dissolved Oxygen, Winkler method	
Range:	0.0 to 10.0 mg/L
Reagents	
HI 93732-01	Reagent kit for 100 tests (O ₂)
HI 93732-03	Reagent kit for 300 tests (O ₂)
Fluoride HR, SPADNS method	
Range:	0.0 to 20.0 mg/L
Reagents	
HI 93739-01	Reagent kit for 100 tests (F ⁻ HR)
HI 93739-03	Reagent kit for 300 tests (F ⁻ HR)
Fluoride LR, SPADNS method	
Range:	0.00 to 2.00 mg/L
Reagents	
HI 93729-01	Reagent kit for 100 tests (F ⁻ LR)
HI 93729-03	Reagent kit for 300 tests (F ⁻ LR)
HI 93703-53	Optional Reagent for chlorine removal

* The sulphuric acid method is officially recognized by EPA for wastewater.

** HI 93754F-25 and HI 93754G-25 are compliant to ISO 15705 method.

Code	Description
Hardness calcium, calmagite method	
Range:	0.00 to 2.70 mg/L
Reagents	
HI 93720-01	Reagent kit for 100 tests (Ca hardness, CaCO ₃)
HI 93720-03	Reagent kit for 300 tests (Ca hardness, CaCO ₃)
Hardness calcium, colorimetric method	
Range:	0 to 500 mg/L
Reagents	
HI 93756-01	Reagent kit for 100 tests (Ca hardness, CaCO ₃)
HI 93756-03	Reagent kit for 300 tests (Ca hardness, CaCO ₃)
Hardness magnesium, EDTA method	
Range:	0.00 to 2.00 mg/L
Reagents	
HI 93719-01	Reagent kit for 100 tests (Mg hardness, CaCO ₃)
HI 93719-03	Reagent kit for 300 tests (Mg hardness, CaCO ₃)
Hardness, total, colorimetric method	
Range:	0 to 750 mg/L
Reagents	
HI 93735-0	Reagent kit for 100 tests (0-750 mg/L CaCO ₃)
HI 93735-00	Reagent kit for 100 tests (0-250 mg/L CaCO ₃)
HI 93735-01	Reagent kit for 100 tests (200-500 mg/L CaCO ₃)
HI 93735-02	Reagent kit for 100 tests (400-750 mg/L CaCO ₃)
Hydrazine, p-dimethylaminobenzaldehyde method	
Range:	0 to 400 µg/L
Reagents	
HI 93704-01	Reagent kit for 100 tests (N ₂ H ₄)
HI 93704-03	Reagent kit for 300 tests (N ₂ H ₄)
Iodine, DPD method	
Range:	0.0 to 12.5 mg/L
Reagents	
HI 93718-01	Reagent kit for 100 tests (I ₂)
HI 93718-03	Reagent kit for 300 tests (I ₂)
Iron HR, phenantroline method	
Range:	0.00 to 5.00 mg/L
Reagents	
HI 93721-01	Reagent kit for 100 tests (Fe HR)
HI 93721-03	Reagent kit for 300 tests (Fe HR)
Iron LR, TPTZ method	
Range:	0 to 400 µg/L
Reagents	
HI 93746-01	Reagent kit for 50 tests (Fe LR)
HI 93746-03	Reagent kit for 150 tests (Fe LR)
Manganese HR, periodate method	
Range:	0.0 to 20.0 mg/L
Reagents	
HI 93709-01	Reagent kit for 100 tests (Mn HR)
HI 93709-03	Reagent kit for 300 tests (Mn HR)
Manganese LR, PAN method	
Range:	0 to 300 µg/L
Reagents	
HI 93748-01	Reagent kit for 50 tests (Mn LR)
HI 93748-03	Reagent kit for 150 tests (Mn LR)

*** It is a recommended method for generic analysis and if chlorine is absent.

**** Some meters allow the user to extend the range to the value between brackets.

Reagents for Photometers

Code	Description
Molybdenum, mercaptoacetic acid method	
Range:	0.0 to 40.0 mg/L
Reagents	
HI 93730-01	Reagent kit for 100 tests (Mo)
HI 93730-03	Reagent kit for 300 tests (Mo)
Nickel HR, fotometric method	
Range:	0.00 to 7.00 mg/L
Reagents	
HI 93726-01	Reagent kit for 100 tests (Ni HR)
HI 93726-03	Reagent kit for 300 tests (Ni HR)
Nickel LR, PAN method	
Range:	0.000 to 1.00 0 mg/L
Reagents	
HI 93740-01	Reagent kit for 50 tests (Ni LR)
HI 93740-03	Reagent kit for 150 tests (Ni LR)
Nitrate, cadmium reduction method	
Range:	0.0 to 30.0 mg/L
Reagents	
HI 93728-01	Reagent kit for 100 tests (NO ₃ ⁻ -N)
HI 93728-03	Reagent kit for 300 tests (NO ₃ ⁻ -N)
Nitrate, cromotropic acid method	
Range:	0.0 to 30.0 mg/L
Reagents	
HI 93766-50	Reagent kit for 50 test vials (NO ₃ ⁻ -N)
Nitrite HR, ferrous sulphate method	
Range:	0 to 150 mg/L
Reagents	
HI 93708-01	Reagent kit for 100 tests (NO ₂ ⁻ HR)
HI 93708-03	Reagent kit for 300 tests (NO ₂ ⁻ HR)
Nitrite LR, diazotation method	
Range:	0.00 to 0.35 mg/L (1.500 mg/L)****
Reagents	
HI 93707-01	Reagent kit for 100 tests (N-NO ₂ ⁻ LR)
HI 93707-03	Reagent kit for 300 tests (N-NO ₂ ⁻ LR)
Nitrogen, total, HR, cromotropic acid method	
Range:	10 to 150 mg/L
Reagents	
HI 93767B-50	Reagent kit for 50 test vials (N)
Nitrogen, total, LR, cromotropic acid method	
Range:	0.0 to 25.0 mg/L
Reagents	
HI 93767A-50	Reagent kit for 50 test vials (N)
Ozone, DPD method	
Range:	0.00 to 2.00 mg/L
Reagents	
HI 93757-01	Reagent kit for 100 tests (O ₃)
HI 93757-03	Reagent kit for 300 tests (O ₃)
HI 93703-52	Optional reagent for 100 tests (glycin)
pH, phenol red method	
Range:	5.9(6.5)**** to 8.5 pH
Reagents	
HI 93710-01	Reagent kit for 100 tests (pH)
HI 93710-03	Reagent kit for 300 tests (pH)
Phosphate HR, amino acid method	
Range:	0.0 to 30.0 mg/L
Reagents	
HI 93717-01	Reagent kit for 100 tests (PO ₄ ³⁻ HR)
HI 93717-03	Reagent kit for 300 tests (PO ₄ ³⁻ HR)

Code	Description
Phosphate LR, ascorbic acid method	
Range:	0.00 to 2.50 mg/L
Reagents	
HI 93713-01	Reagent kit for 100 tests (PO ₄ ³⁻ LR)
HI 93713-03	Reagent kit for 300 tests (PO ₄ ³⁻ LR)
Phosphorus acid-hydrolyzable, ascorbic acid method	
Range:	0.00 to 5.00 mg/L
Reagents	
HI 93758B-50	Reagent kit for 50 test vials (PO ₄ ³⁻)
Phosphorus, amino acid method	
Range:	0.0 to 15.0 mg/L
Reagents	
HI 93706-01	Reagent kit for 100 tests (P)
HI 93706-03	Reagent kit for 300 tests (P)
Phosphorus, reactive, ascorbic acid method	
Range:	0.00 to 5.00 mg/L
Reagents	
HI 93758A-50	Reagent kit for 50 test vials (PO ₄ ³⁻)
Phosphorus reactive HR, vanadomolibdophosphoric acid method	
Range:	0.0 to 100.0 mg/L
Reagents	
HI 93763A-50	Reagent kit for 50 test vials (PO ₄ ³⁻)
Phosphorus total, ascorbic acid method	
Range:	0.00 to 3.50 mg/L
Reagents	
HI 93758C-50	Reagent kit for 50 test vials (PO ₄ ³⁻)
Phosphorus total, HR, vanadomolibdophosphoric acid method	
Range:	0.0 to 100.0 mg/L
Reagents	
HI 93763B-50	Reagent kit for 50 test vials (PO ₄ ³⁻)
Potassium, turbidimetric method	
Range:	0.0 to 50.0 mg/L
Reagents	
HI 93750-01	Reagent kit for 100 tests (K)
HI 93750-03	Reagent kit for 300 tests (K)
Silica, LR heteropoly blue method	
Range:	0.00 to 2.00 mg/L
Reagents	
HI 93705-01	Reagent kit for 100 tests (SiO ₂)
HI 93705-03	Reagent kit for 300 tests (SiO ₂)
Silver, PAN method	
Range:	0.000 to 1.000 mg/L
Reagents	
HI 93737-01	Reagent kit for 50 tests (Ag)
HI 93737-03	Reagent kit for 150 tests (Ag)
Sulphate, turbidimetric method	
Range:	0 to 150 mg/L
Reagents	
HI 93751-01	Reagent kit for 100 tests (SO ₄ ²⁻)
HI 93751-03	Reagent kit for 300 tests (SO ₄ ²⁻)
Zinc, zincon method	
Range:	0.00 to 3.00 mg/L
Reagents	
HI 93731-01	Reagent kit for 100 tests (Zn)
HI 93731-03	Reagent kit for 300 tests (Zn)

* The sulphuric acid method is officially recognized by EPA for wastewater.

** HI 93754F-25 and HI 93754G-25 are compliant to ISO 15705 method.

*** It is a recommended method for generic analysis and if chlorine is absent.

**** Some meters allow the user to extend the range to the value between brackets.